### **XUND Solutions GmbH**

Dorotheergasse 10/12a A-1010 Vienna Tel. +43 699 1900 24 93

Web: <u>www.xund.ai</u> SRN: AT-MF-000014461



# **Declaration of Conformity**

## **Product details:**

Product name	XUND v1	
Software version	v1.24	
Basic UDI-DI	1918000298XUND01HA	
Classification according 2017/745	Class IIa ; according to rule 11	

### Assessment details:

Notified body	TÜV SÜD Product Service GmbH (0123)	
Route of directive regulation 2017/745	according to Annex IX	
Certificates	G10 111420 0001	

## **Used standards:**

Harmonized standards and other standards					
Standard	No.	Part	Date / Release	Title	
ISO	20417	n/a	2021	Information supplied by the manufacturer of medical devices	
EN ISO	13485	n/a	2016+A11/2021	Medical devices - Quality management systems - Requirements for regulatory purposes	
EN ISO	14971	n/a	2019+A11/2021	Medical devices - Application of risk management to medical devices	
EN ISO	15223	1	2016	Medical devices - Symbols to be used with medical device labels, labeling and information to be supplied - Part 1: General requirements	
EN IEC	62304	n/a	IEC 62304:2006 + A1:2015); German Version EN 62304:2006 + Cor.:2008 + A1:2015 DRAFT: IEC/DIS 62304:2019(E)	Medical device software - Software life-cycle processes	
IEC	62366	1	2015	Medical devices - Part 1: Application of usability engineering to medical devices	
IEC	82304	1	2016	Health software — Part 1: General requirements for product safety	
EEC	MEDDEV 2.7/1 rev. 4	4	2016	CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES UNDER DIRECTIVES 93/42/EEC and 90/385/EEC	

We declare under sole responsibility that the products described above as delivered are in compliance with the regulation 2017/745. The products are CE marked.



Vienna, 11.04.2024



Tamás Petrovics Co-Founder & CEO